

Calibration certificate

Model: Current measuring clamp
A 1019

Serial No.: 14210576

Date: 12.8.2014

Performed By: Mihael Šraj

Date Placed In Service: _____ Due Date: _____*

Metrel Recommended Cal Interval: 12 months

* The due date may be established (by the customer) by adding the "Recommended Cal Interval" to the "Date Placed In Service."

No.	Function	Input	Low limit	Output	Uncertainty	High limit
1	Outlook			PASS		
2	Range 1000 A / 1 A	100 mA	97,2 uA	99.6 uA	0,7 uA	102,8 uA
	Frequency 50 Hz	10 A	9,77 mA	10.01 mA	0,04 mA	10,23 mA
		50 A	49,1 mA	49.8 mA	0,1 mA	50,9 mA
		200 A	197,6 mA	200.4 mA	0,6 mA	202,4 mA
		1000 A	990 mA	999 mA	3 mA	1010 mA
		1200 A	1188 mA	1199 mA	4 mA	1212 mA

All results in accordance with technical specification.

Reference instruments:

No.	Instrument	Type	Certificate No.	Due
1	Calibration System	9100, Wavetek	14C00944	31.7.2015



Measurement and Regulation Equipment Manufacturers

METREL d.d.
Ljubljanska c. 77
SI - 1354 HORJUL
SLOVENIA

Tel.: (+386 1) 7558 200
Fax.: (+386 1) 7549 095
<http://www.metrel.si>
e-mail: metrel@metrel.si